

**Questions and Answers for Solicitation # - NNC10340422Q – 30 KeV High Precision
RHEED – System with differential pump**

1) Q - Do you want the vacuum pressure shut-off?

A – Yes

2) Q – Do you want the pumping package included?

A - Main pumping package should be included. Second pumping package for high pressure measurements should be just an option.

3) Q – The solicitation requests "• Electronic control to change beam incidence angle without changing spot position". I don't know if Beam rocking should be included in the basic package or as an option in my response.

A – Should be an option not in the basic package.